

EC CERTIFICATE

Full Quality Assurance System

Certificate No.:
241154-2017-CE-ITA-NA-PS Rev. 3.0

Project No.:
PRJC-403782-2012-MSL-ITA

Valid Until:
27 May 2024

This is to certify that the quality system of:

ANGELANTONI LIFE SCIENCE S.r.l.

Località Cimacolle, 464 06056 MASSA MARTANA Italy

For design, production and final product inspection/testing of:
Human Tissues Cooling Units

Has been assessed with respect to:

**THE CONFORMITY ASSESSMENT PROCEDURE DESCRIBED IN
ANNEX II EXCLUDING SECTION 4 OF COUNCIL DIRECTIVE
93/42/EEC ON MEDICAL DEVICES, AS AMENDED**

and found to comply.

Further details of the product(s) and conditions for certification are given overleaf.

Place and date:
Høvik, 19 December 2019

For:
DNV GL PRESAFE AS
Notified Body No.: 2460



Palani Damodharan

The certificate is digitally verified by blockchain technology. For more info, see www.dnvgl.com/assurance/certificates-in-the-blockchain.html



Certificate No.:
241154-2017-CE-ITA-NA-PS Rev. 3.0

Project No.:
PRJC-403782-2012-MSL-ITA

Valid Until:
27 May 2024

Jurisdiction

Application of Council Directive 93/42/EEC of 14 June 1993, adopted as "Forskrift om Medisinsk Utstyr" by the Norwegian Ministry of Health and Care Services.

Certificate history:

Revision	Description	Issue Date
0.0	Original Certificate - Replaces Certificate 66510-2009-CE-ITA-DNV Rev. 5.0 (NB 0434), following transfer of Notified Body functions to DNV GL NEMKO Presafe AS (NB 2460)	2017-10-30
1.0	Blood Banks: - Deletion of discontinued Devices - Hemonine, Hemosafe, Plasmafrost; Refrigeration Chambers (all); Freezer Chambers (all) ; - New device added: Hemosafe 2.0 - Unmerging Blast Freezers from Freezers	2018-10-11
2.0	Blood Banks: - New devices added Emobanks & Plasmabanks (EB&PB) - Deletion of discontinued Devices - MINI, BBR	2019-11-19
3.0	Recertification Freezer - New devices added Nexus Evo, Smart Freezer Deletion of discontinued Devices - Platinum, Iridium, PlatinumNext, IridiumNext, Kryos MD	2019-12-19

Products covered by this Certificate:

Product Description	Product Name	Class
Blood and Blood Cells Cooling Units	Blood Banks	IIa
Human Tissues (Blood, blood cells, tissues and fluids) Cooling Units	Freezers	IIa
Human Tissues (Plasma) Fast Cooling Chambers	Blast Freezers	IIa

The complete list of devices is filed with the Notified Body

Sites covered by this certificate

Site Name	Address
ANGELANTONI LIFE SCIENCE S.r.l.	Località Cimacolle, 464 06056 MASSA MARTANA

Certificate No.:
241154-2017-CE-ITA-NA-PS Rev. 3.0

Project No.:
PRJC-403782-2012-MSL-ITA

Valid Until:
27 May 2024

EU Representative

N.A.

Terms and conditions

The certificate is subject to the following terms and conditions:

- Any producer (see 2001/95/EC for a precise definition) is liable for damage caused by a defect in his product(s), in accordance with directive 85/374/EEC, as amended, concerning liability of defective products.
- The certificate is only valid for the products and/or manufacturing premises listed above.
- The Manufacturer shall fulfil the obligations arising out of the quality system as approved and uphold it so that it remains adequate and efficient.
- The Manufacturer shall inform Presafe of any intended updating of the quality system and Presafe will assess the changes and decide if the certificate remains valid.
- Periodical audits will be held, in order to verify that the Manufacturer maintains and applies the quality system. Presafe reserves the right, on a spot basis or based on suspicion, to pay unannounced visits.

The following may render this Certificate invalid:

- Changes in the quality system affecting production.
- Periodical audits not held within the allowed time window.

Conformity declaration and marking of product

When meeting with the terms and conditions above, the producer may draw up an EC declaration of conformity and legally affix the CE mark followed by the Notified Body identification number of Presafe.

End of Certificate